



drexel® electronics & engineering products (p) ltd

Works: 442, JLPL, Industrial Area
Sector 82, Mohali (PB) 160055
Ph: +91172 4190098
Mob: +91 9988610098
Email: sales@drexel.in
Web: www.drexel.in

Corp: 13/2, 2nd & 3rd Floor
Opp. Bus terminal Mehrauli,
New Delhi 110030
Toll Free: 1800 123440098
Mob: +91 9316660098
info@drexel.in

Calibration Certificate

Sample Code	DXL/2026/12403		
Customer Name & Address:	M/s Haat Supreme Wastech Private Limited #V.P.O. Bazida Jatan, Karnal, Haryana-132037		
Calibration Date:	17.02.2026	Calibration Due Date:	16.08.2026

PRODUCT DETAIL:

Product Name:	Online Emission Monitoring System	Make & Model:	Drexel & DX4B
Product Sr. No.:	117	Voltage & Frequency:	220V AC/50 Hz

Results:

Sr. No.	Sensor Applicable	Gas Conc.	Current (Range 4 mA – 20 mA)			
			Minimum	Span	Maximum	Remark
1	CO	500 PPM	0 PPM	491 PPM	10,000 PPM	ok
2	CO2	7%	0 PPM	6.1%	11	ok
3	HCL	500 PPM	0 PPM	493 PPM	1000 PPM	ok
4	NOX	500 PPM	0 PPM	495 PPM	1000 PPM	ok
5	SPM	Zero Check			Isokinetic Technique	
	Suction Unit:					
1.	Suction Pump	ok	Auto- clean Filter Assembly			ok
2.	Filter Assembly	ok	Purging Cycle			ok

Remarks: System is completely tested & inspected properly and found OK.

Israfil

Tested By
Mr. Israfil
Field Engineer



Mr. Gurpreet Singh Field

Checked By
Mr. Gurpreet Singh Field
Senior Engineer (Technical)

